

## **Product Datasheet**

## Mouse T4(Thyroxine) ELISA Kit (orb1817324)

**Description** This assay employs the competitive inhibition enzyme immunoassay technique.

The microtiter plate provided in this kit has been pre-coated with Mouse T4 protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse T4. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$  10nm. The concentration of Mouse T4 in the samples is then determined by comparing

the OD of the samples to the standard curve.

**Reactivity** Mouse

**Range** 4.69-300 ng/mL

**Concentration** 300 ng/mL

**Note** For research use only

**Application notes** standard: 300 ng/mL. Test principle: This assay employs the competitive

inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Mouse T4 protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse T4. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of  $450 \, \mathrm{nm} \pm 10 \, \mathrm{nm}$ . The concentration of Mouse T4 in the samples is then determined by comparing the OD of the samples to the standard curve


**Sample Types** serum, plasma and other biological fluids

**Sensitivity** 1.51 ng/mL

**Expiration Date** Please enquire.







68 TW Alexander Drive, Durham, NC, 27713, United States

Email: <u>info@biorbyt.com</u>, <u>support@biorbyt.com</u>
Phone: <u>+1 (415) 906-5211</u> | Fax: <u>+1 (415) 651-8558</u>