

Product Datasheet

Human Metallothionein 2 (MT2) ELISA Kit (orb1948838)

Description Human Metallothionein 2 (MT2) ELISA Kit

Reactivity Human

Range 0.32-20ng/mL

Target MT2

Note For research use only

Application notes This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided

in this kit has been pre-coated with an antibody specific to Human MT2.

Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human MT2 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that

contain Human MT2, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The OD value is proportional to the concentration of Human MT2. You can calculate the

concentration of Human MT2 in the samples by comparing the OD of the

samples to the standard curve.

Sample Types Serum, Plasma, Tissue homogenate and Other biological samples


Uniprot ID P02795

Sensitivity 0.19ng/mL

Expiration Date Please enquire.







Email: <u>info@biorbyt.com</u>, <u>support@biorbyt.com</u>
Phone: <u>+1 (415) 906-5211</u> | Fax: <u>+1 (415) 651-8558</u>