

Product Datasheet

Human GSTω1(Glutathione S Transferase Omega 1) ELISA Kit (orb1949034)



www.biorbyt.com

Descriptionnts. Human GSTω1(Glutathione S Transferase

Reactivity Human

Range 0.16-10ng/mL

Target GST ω 1

Note For research use only

Application notes This ELISA kit uses the Sandwich-ELISA

principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human GST ω 1. Standards or samples are added to the micro ELISA plate wells and combined

with the specific antibody. Then a biotinylated detection antibody specific for Human GST ω 1 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human GST ω 1, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzymesubstrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a

wavelength of 450 nm \pm 2 nm. The OD value is proportional to the concentration of Human GST ω 1. You can calculate the concentration of Human GST ω 1 in the samples by comparing the OD of the

samples to the standard curve.

Sample Types serum, plasma, Tissue homogenate and

Other biological samples

Uniprot ID P78417

Sensitivity 0.1 ng/mL

Expiration Date Please enquire.

