

Product Datasheet

HighFidelity ORANGE PCR Labeling Testkit (orb692242)

Description	HighFidelity ORANGE PCR Labeling Testkit is designed to produce randomly AF594-, ATTO594-and TexasRe...
Storage	store at -20 °C
Note	For research use only
Application notes	Spectroscopic Propertie: AF594: λ_{exc} 590 nm, λ_{em} 617 nm, ϵ 92.0 L mmol ⁻¹ cm ⁻¹ (Tris-HCl pH 7.5), ATTO594: λ_{exc} 602 nm, λ_{em} 626 nm, ϵ 120.0 L mmol ⁻¹ cm ⁻¹ (Tris-HCl pH 7.5), TexasRed: λ_{exc} 583 nm, λ_{em} 603 nm, ϵ 112.0 L mmol ⁻¹ cm ⁻¹ (Tris-HCl pH 7.5).
Hazard Information	Non-Toxic
Expiration Date	12 months from date of receipt.